



# DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

## EU Representative

SUNGO Europe B.V.  
Olympisch Stadion 24, 1076DE  
Amsterdam, Netherlands  
SRN: NL-AR-000000247

## Conformity Assessment

**Conformity Assessment Procedure**  
Annex II+III of Regulation (EU) 2017/745

**Applicable Standards**  
EN ISO 14971: 2019  
EN ISO 15223-1: 2016  
EN 1041:2008+A1:2013  
ISO 10993-1: 2018  
EN ISO 10993-5: 2009  
EN ISO 10993-10: 2013

### Remark

*The declaration of conformity is valid in connection with the release technical document CE/MDR-CADB-02.*

*All the supporting documentation is retained at the premises of the manufacturer.*

*The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.*

## Manufacturer

**Name:** SHANGHAI WILSON PRECISION TECHNOLOGY CO., LTD.

**Address:** Area A, B, C of 1F, Area D, E of 2F), Building 8, No.868, Zhenchen Road, Baoshan District, Shanghai, China.

## Product Information

**Name:** Tungsten Carbide Burs For Dental Trimming

**Model:** XXC、XC、C、M、F、XF、XXF


**GMDN:** 16668

**Basic UDI-DI:** /

**Classification:** Class I, According to Rule 1, Annex VIII, Regulation (EU) 2017/745

## Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature: 

Date: 2021.07.08

Position: GM

Place: Shanghai/China

